

Early and Immediately Restored and Loaded Dental Implants for Single-Tooth and Partial-Arch Applications

Jeffrey Ganeles, DMD¹/Daniel Wismeijer, DDS, PhD²

Purpose: The objective of this consensus committee report was to review the available literature published predominantly in refereed journals to summarize findings, data, and conclusions as they related to reduced healing times and protocols for single-tooth and partial-arch clinical situations. Early loading of dental implants has been defined as restoration of implants in or out of occlusion at least 48 hours after implant placement, but at a shorter time interval than conventional healing. Immediate loading or restoration has been defined as attachment of a restoration in or out of direct occlusal function within 48 hours of surgical placement. **Materials and Methods:** Six articles addressing early loading, with a mixture of single-tooth and partial-arch clinical conditions and including some controlled cohort studies, were reviewed. Immediate loading or restoration of dental implants in single-tooth and partial-arch applications, was extensively reviewed. An attempt was made to isolate and categorize similar case types to discern trends and relevant factors. Variables that were considered included single- or multiple-tooth conditions, immediate or delayed placement in extraction sockets, effect of implant surface and geometry, bone quality, implant stability, surgical technique, occlusal design, effect of cigarette smoking, and stability of results. **Results:** Combined data from 6 early loading studies on single-tooth and partial-arch applications revealed 1,046 implants with a survival rate of 98.2%. Long-term data for most of the early loading studies were not yet available. Most of the publications on immediate loading or restoration of dental implants were written as case series rather than scientific studies. **Discussion and Conclusions:** In general, most publications indicated that with attention to appropriate factors, implant survival with immediate restoration was comparable to the results with conventional and early loading protocols. It should be recognized that, with few exceptions, these conclusions may be misleading statistical phenomena of the authors, as most publications were written by exceptionally experienced, highly skilled practitioners working under tightly controlled clinical conditions on a relatively small, statistically inconclusive number of implants and patients. INT J ORAL MAXILLOFAC IMPLANTS 2004;19(SUPPL):92–102

Key words: dental implants, early loading, fixed partial denture, immediate function, immediate loading, immediate restoration, provisional denture

A far greater number of patients are edentulous in a single-tooth gap or partial-arch space than are completely edentulous. The opportunity to provide

implant-supported tooth replacement for these patients significantly exceeds the opportunity for those who are completely edentulous.¹ The biomechanics of implants in these situations are significantly different than in completely edentulous conditions, particularly in the context of immediate restoration of these implants. Abundant evidence clearly exists to support immediate loading of implants under full-arch clinical conditions. Limiting implant micromotion below the threshold that could interfere with osseointegration, despite occlusal function, has been well documented and elucidated in the previous section and by many authors.^{2–4} Methods to achieve this objective include placing an adequate number of (usually) threaded implants into sufficiently dense bone. Stiff restorative materials are

¹Periodontist, Florida Institute for Periodontics and Dental Implants, Boca Raton, Florida; Adjunct Associate Professor, Nova Southeastern University, College of Dental Medicine, Ft Lauderdale, Florida.

²Oral Implantologist and Maxillofacial Prosthodontist, Department of Special Dental Care and Maxillofacial Prosthodontics, Amphia Teaching Hospital, Bredda, The Netherlands; Referral Practice for Oral Implantology, Veluwezoom, Dieren, The Netherlands.

Correspondence to: Dr Jeffrey Ganeles, Florida Institute for Periodontics and Dental Implants, Boca Raton, FL 33431. Fax: +561-912-9883. E-mail: jganeles@perio-implant.com

used to splint implants together using the principles of cross-arch stabilization to distribute occlusal forces between the implants and immobilize them during patient function.

This article will attempt to summarize and organize the relevant literature and factors that pertain to immediate and early restoration and loading of implants in single-tooth and partial-arch applications.

METHODS

A review of the available literature from a MEDLINE search and manual journal searches revealed numerous strategies for achieving osseointegration in single-tooth and partial-arch clinical conditions. It should be noted that there is comparatively less information available on these conditions than for full-arch rehabilitation. The published information shows a tendency toward case series and case studies rather than controlled studies in this area. Further, because many publications report on case series, different clinical factors occur simultaneously. For example, authors such as Degidi and Piattelli⁵ and Glauser and coworkers⁶ report on series of patients with single and multiple implants in healed alveolar ridges and extraction sockets, who may or may not smoke, using different implant types or surfaces.

Drawing solid conclusions from data like these is hindered by the introduction of confounding or conflicting variables. Yet it is possible to recognize clinical principles, relevant trends, strategies, and insights from examining these publications.

EARLY RESTORATION AND LOADING

This ITI Consensus Conference Introductory Section has previously defined loading protocols and definitions. Prosthetic connection in occlusion to an implant within 48 hours of surgical implant placement is considered *immediate loading*. *Conventional loading* has been defined as restoration and loading of an implant after a healing period of 3 to 6 months. *Early loading* has been defined as prosthetic loading or utilization of an implant at any time period between immediate and conventional loading.

Cochran and associates⁷ reported on a longitudinal, prospective, multicenter study of early loading of 383 ITI SLA implants (Institut Straumann, Waldenburg, Switzerland) placed in the posterior jaws of 307 patients. The implants were allowed to heal for 42 to 63 days in classes 1 to 3 bone and for 84 to 105 days in class 4 bone prior to restoration.⁸ Patients who were heavy smokers or who had inad-

equate bone volume, bruxism, or immediate placement indications were excluded. At abutment placement and torque application, 3 implants were mobile and removed, while 3 rotated and 6 were associated with pain. All implants associated with either pain or rotation were allowed additional healing time and eventually became clinically integrated and were restored, resulting in a survival rate of 99.1%. Three hundred twenty-six implants had passed the 1-year evaluation period and 138 had passed the 2-year period without additional changes in clinical parameters.

Rocuzzo and colleagues⁹ reported on a prospective, split-mouth design study comparing early loading of 68 SLA implants (sandblasted, large-grit, acid-etched) restored at 6 weeks and 68 identically shaped titanium plasma-spray (TPS) surface implants restored at 12 weeks in 32 healthy patients (all implants ITI/Institut Straumann). Solid restorative abutments were torqued to 35 Ncm at the time of restoration. Four of 68 test SLA implants rotated and the patients experienced pain at the 6-week abutment placement procedure; the implants were allowed to heal an additional 6 weeks before re-torquing. None of the control implants demonstrated complications at restoration. After a 1-year evaluation, the authors noted 100% success with no significant differences in clinical parameters between the 2 groups of implants, including radiographic evaluation.

In another prospective study on ITI implants, Rocuzzo and Wilson¹⁰ reported on 36 maxillary posterior implants placed in 19 nonsmoking patients using an altered surgical protocol to increase initial implant stability. Minimal drilling was performed, in favor of bone condensation, to compact and compress maxillary trabecular bone during implant placement. Abutments were torqued to 15 Ncm after 43 days, and the implants were restored with provisional restorations in infraocclusion. After an additional 6 weeks, the abutments were torqued to 35 Ncm for definitive restoration fabrication. One implant rotated with pain at 42 days and was subsequently removed. The other 35 implants were restored uneventfully, leading to a 1-year survival rate of 97.2%. The authors reported implant clinical indices similar to the 6-week period, although marginal bone loss of 0.55 ± 0.49 mm versus the immediate postoperative radiographs was noted.

Testori and coworkers¹¹ reported on a longitudinal, prospective, multicenter early loading study of 475 Osseotite implants (3i/Implant Innovations, West Palm Beach, FL) in posterior sextants of 175 patients restored at 2 months. Patients who were

bruxers or had periodontal or systemic diseases were excluded, while smokers were not. Six of 475 implants failed to integrate within the first 2 months and were considered early failures, while 3 failed after restoration and were considered late failures. The cumulative survival rate was 97.7% after 3 years.

Bogaerde and colleagues¹² reported a prospective study of 31 nonsmoking, nonbruxing patients with 36 edentulous areas treated with 124 Brånemark System machined-surface Mk IV implants (Nobel Biocare, Göteborg, Sweden) provisionally restored 7 to 20 days after surgical placement. One hundred one of the implants were placed in partial-arch applications. One of the inclusion criteria was the ability to achieve 40 Ncm of insertion torque at implant placement, which was generally achieved by underpreparation of the diameter of the osteotomies. Provisional restorations with light occlusal contact were placed at a mean of 11 days postsurgically (maximum of 20 days postsurgically). Ninety-seven of 101 (96%) implants in partial arches integrated, with 3 early failures and 1 late loss at 6 months. Clinical and radiographic evaluation appeared to indicate stable results at 18 months, although according to the authors, many of the radiographic data were not readable or usable for analysis.

Cooper and coworkers¹³ reported on 47 patients with 53 early loaded 11- to 17-mm Astra Tech ST implants (Astra, Mölndal, Sweden) to replace 53 maxillary anterior single teeth. Patients were excluded from treatment if they were positive for bruxism, unstable posterior occlusion, daily cigarette smoking, uncontrolled periodontal disease, systemic disease, or mobility of the teeth adjacent to the planned implant site. Acrylic resin restorations were placed into occlusal contact 3 weeks after surgery, at which time abutments were torqued with hand pressure. After 8 weeks, final abutments were torqued to 20 Ncm and definitive restorations were placed. Two implants failed following provisional restoration placement, while 51 integrated, resulting in an implant survival rate of 96.2%.

Drawing conclusions from the limited literature on early loading in partial-arch applications is difficult because of the paucity of information. Table 1 summarizes the information available from the 6 studies that were reviewed, although it should be recognized that all of the articles are not directly comparable. These reports indicate that early loading of 1,046 implants in 611 patients resulted in survival or success of 1,027 implants, for a mean survival rate of 98.2%. All authors indicated high success rates of implants and restorations consistent with delayed loading protocols, but few long-term

data have yet been published. Common strategies used by most of the authors, with the exception of Bogaerde and associates,¹² appear to include rough-surfaced implants, infraocclusion, and enhanced surgical stability. None of the authors of the articles reviewed reported placement of implants in immediate extraction sockets with this loading protocol.

IMMEDIATE RESTORATION AND LOADING

Early publications on immediate restoration of single, unsplinted implants in the esthetic zone were presented as case reports and series. Kupeyan and May¹⁴ and Wöhrle¹⁵ reported on series of 10 and 14 immediately restored implants, respectively, in the maxillary anterior region. Kupeyan and Kay performed their study in healed ridges with machined titanium Brånemark System implants (Nobel Biocare), while Wöhrle reported on roughened-surface Steri-Oss Replace implants (Nobel Biocare) in immediate extraction sites. Both groups indicated that all implants clinically integrated and remained stable for the observation periods of 6 months to 3 years.

Additional case reports of small series of patients by Andersen and coworkers,¹⁶ Aires and Berger,¹⁷ Touati and Guez,¹⁸ Lorenzoni and coworkers,¹⁹ Kan and associates,²⁰ and Cannizzaro and Leone²¹ confirmed the observations of 100% survival of single-tooth replacement in the maxillary anterior region. All authors advocated maximization of implant stability by using long implants and eliminating occlusal contact in centric and excursive movements. Lorenzoni and coworkers¹⁹ advocated the use of an occlusal splint for 8 weeks to prevent loading of the restoration by nonocclusal forces such as the tongue or food bolus. Kan and associates²⁰ placed patients on a liquid diet for 2 weeks postoperatively, followed by a soft diet for 5 months. With the exception of Andersen and colleagues,¹⁶ who indicated that 2 of their 8 patients were cigarette smokers, it appears that patients who smoked more than 10 cigarettes per day or had parafunctional occlusal habits such as bruxism or clenching were excluded from treatment by most authors.

Ericsson and associates²² reported on 14 consecutive patients treated with Brånemark System MKII implants (Nobel Biocare) in the maxillary anterior area. Nonsmoking patients with negative histories for parafunctional habits had implants placed in healed ridges and immediately restored out of occlusion. Two (14%) implants failed to integrate within the first 5 months. All others were clinically integrated and maintained stable radiographic bone levels throughout the observation period of 18 months.

Table 1 Publications on Early Loading

| Authors | Type of study | Implant system | No. of patients | Groups | No. of implants | No. of early loaded implants | Time before loading | Survival/success | Other |
|------------------------------------|---------------|---------------------------|-----------------|--|---------------------------------|------------------------------|---|------------------|---|
| Roccuzzo et al 2001 ⁹ | Pros sm | Straumann SLA versus TPS | 32 | 68 pairs of implants | 68 SLA versus 68 TPS; 136 total | 68 | Provisional restoration: 6 weeks SLA, 12 weeks TPS | 68 (100%) | 4 SLA implants rotated at 6 weeks, but subsequently healed |
| Testori et al 2002 ¹¹ | Pros | Osseotite (3i) | 175 | 99 single-tooth, 119 short-span prostheses, 11 full-arch | 405 | 405 | Provisional restoration: 2 months | 396 (97.8%) | 97.8% success |
| Cooper et al 2001 ¹³ | Pros | Astra | 47 | 53 single-tooth implants | 53 | 53 | Provisional restoration after 3 weeks, definitive 8 to 12 weeks | 51 (96.2%) | 2 implants lost |
| Cochran et al 2002 ⁷ | Pros | Straumann SLA | 307 | Various | 383 | 383 | Provisional restoration: 6 weeks in type 1 to 3 bone, 12 to 15 weeks in type 4 bone | 380 (99.1%) | 3 rotated and were removed, 6 others painful at evaluation subsequently healed; 3 failed of 383 |
| Roccuzzo/Wilson 2002 ¹⁰ | Pros | Straumann SLA | 19 | Various in type 4 maxillary bone using bone condensation | 36 | 36 | 6 weeks for provisional with 15 Ncm torque, 12 weeks final torque | 35 (97.1%) | 97.1% survival |
| Bogaerde et al 2003 ¹² | Pros | Brånemark System machined | 31 | 56 implants in mandible, 45 implants in maxilla | 101 | 101 | Provisional restoration: 7 to 20 days | 97 (96.0%) | 96% success |
| Totals | | | 611 | | 1,046 early loaded implants | | | | 1,027 (98.2%) surviving/successful implants |

Pros = prospective; sm = split-mouth.

Table 2 Publications on Single-Tooth Immediate Restorations

| Authors | Type of study | Implant system/surface | No. of patients | No. of implants | Successful implants | Success rate (%) |
|---------------------------------------|---------------|------------------------------|---|-----------------|---------------------|------------------|
| Wöhrle 1998 ¹⁵ | Pros | Steri-Oss TPS and HA | Single tooth, 14 | 14 | 14 | 100.0 |
| Kupeyan/May 1998 ¹⁴ | Pros | Brånemark machined | Single tooth, 10 | 10 | 10 | 100.0 |
| Ericsson et al 2000 ²² | Pros | Brånemark machined | 14 | 14 | 12 | 85.7 |
| Hui et al 2001 ²³ | Pros | Brånemark machined | 13 | 13 | 13 | 100.0 |
| Andersen et al 2002 ¹⁶ | Pros | ITI TPS | Single tooth, 8 | 8 | 8 | 100.0 |
| Rocci et al 2003 ²⁶ | Pros | Brånemark machined | Not specified | 27 | 22 | 81.5 |
| Calandriello et al 2003 ²⁴ | Pros | Nobel Biocare TiUnite 5.0 mm | Mandibular molars, 44 | 50 | 50 | 100.0 |
| Lorenzoni et al 2003 ¹⁹ | Pros | Frialit-2 | Single tooth, immediate provisional, 9 | 12 maxillary | 12 | 100.0 |
| Kan et al 2003 ²⁰ | Pros | Steri-Oss Replace | Single tooth, immediate provisional, 35 | 35 maxillary | 35 | 100.0 |
| Cannizzaro/Leone 2003 ²¹ | Rand | Centerpulse Spline twist | Single tooth (immediate loading vs conventional), 28 (2 × 14) | 46 | 46 | 100.0 |
| Degidi/Piattelli 2003 ⁵ | Retro | Multiple | Not specified | 58 | 56 | 96.6 |
| Totals | | | | 287 | 278 | 96.7 |

Pros = prospective; Rand = randomized; Retro = retrospective.

Hui and coworkers²³ studied 2 groups of patients with 24 implants to compare results between immediate placement of implants in 11 extraction sites and immediate placement and restoration in 13 extraction sites in the maxillary anterior region. Heavy smokers and patients with bruxism were excluded. Machined-surface Brånemark System implants 13 to 18 mm long were placed with torque values of 40 to 50 Ncm, with the authors attempting to achieve bicortical anchorage. Provisional restorations were placed the day of surgery with a design of “protected occlusion,” where implants were placed out of contact in all excursive movements. No implants were lost and no complications were encountered. The authors noted that the esthetic outcome of the immediate provisionalization group was better because the provisional restorations preserved the gingival contours.

In an article focusing exclusively on mandibular molars, Calandriello and colleagues²⁴ reported on 44 patients, including 7 smokers, who received fifty 5-mm-wide Nobel Biocare TiUnite implants at least 10 mm in length. All implants were placed in alveolar ridges that had healed for at least 4 months following tooth extraction. They found 100% implant survival at 1 year in bone quality of types 2 and 3.⁸ Despite their restorative protocol of keeping provisional restorations out of occlusion, they noted that several provisional restorations fractured, indicating that some occlusal function occurred.

Cannizzaro and Leone²¹ reported on a prospective study of 28 patients that compared immediate

loading of 46 single implants and 46 matched conventionally loaded implants. All implants were microtextured, self-tapping Centerpulse Spline Twist MTX implants (Centerpulse Dental, Carlsbad, CA) with at least 3.75-mm diameter and 13-mm length. The authors reported a 100% success rate (46 of 46) with the immediately loaded implants and a 97.8% success rate (45 of 46) in the conventionally loaded group. This study is noteworthy for the randomization of other variables, including medical compromise, cigarette smoking, and implant location in patients. Each of the groups of 14 patients included 3 moderate smokers, 1 patient with cardiac disease, 1 patient with controlled hypertension, 1 patient with controlled type 2 diabetes, and 1 patient with asymptomatic HIV infection.

Additional reports of single immediately restored implants are contained within the data from other publications of immediately restored or loaded implants. Table 2 presents the results from 11 publications that include data on single-tooth immediate restoration cases. The listed studies, though not directly comparable, include the observation that 278 of 287 implants achieved clinical osseointegration, for a survival or success rate of 96.7% by various criteria, under immediate restoration conditions. Common themes of the authors include maximization of implant stability and elimination of direct occlusal contact.

Glauser and associates⁶ sought to test the limits of immediate loading, placing 127 consecutive implants (76 maxillary and 51 mandibular) in 41

patients, including smokers. The clinical conditions included single-tooth, partial-arch, and full-arch situations in healed ridges and extraction sockets. Patients with bruxism and imperfect alveolar ridges were not excluded. Brånemark System machined Mk IV implants (Nobel Biocare), with a modest taper, were used to increase stability at the time of surgical placement. Restorations were usually placed the day of surgery and were fabricated in centric occlusal contact without excursive contact. After 1 year, results indicated that 22 implants were lost in 13 patients, including 7 maxillary implants in 1 patient, for a survival rate of 82.7%. Thirty-four percent of 41 implants in the maxillary posterior area failed, while only 9% of the other 86 implants in all other areas failed. Patients with parafunctional habits (22 implants) had failure more often (41%) than nonbruxers (105 implants, or 12%). The authors observed that implants placed in conjunction with guided bone regeneration procedures to cover exposed threads had a better survival rate (90% of 84 implants) than implants placed into adequate ridges (67% of 43 implants). Further, they noted that implants placed into immediate extraction sockets were more successful (44 of 49; 90%) than those placed into healed sites (61 of 78; 78%).

VARIABLES

Extraction Sockets

Malo and coworkers²⁵ placed 94 consecutive machined-surface Brånemark System Mk II implants (Nobel Biocare) in maxillary anterior areas of 49 non-smoking, nonbruxing patients, with 23 areas restored with fixed partial dentures and 31 single-tooth restorations. Fourteen of 57 maxillary and 13 of 37 mandibular implants were placed in fresh extraction sockets. Stability of the implants was enhanced by underdrilling the apical extent of the osteotomies to increase compression of apical bone during implant placement. Four implants placed into immediate extraction sockets failed to integrate, resulting in a success rate of 85.2% in immediate extraction sockets. All other implants achieved clinical integration. Although the protocol called for fabrication of provisional restorations out of occlusion, 12 provisional crowns loosened and three fractured, indicating that occlusal loading occurred during function.

Rocci and associates²⁶ placed 97 machined-surface Brånemark System Mk IV implants (Nobel Biocare) in the partially edentulous maxillary arches of 46 patients, 8 of whom were smokers. Bruxers were excluded. The authors used an elaborate surgical guide and flapless surgery and placed prefabri-

cated provisional restorations. There was no discussion of occlusal design. Eight of 97 (8%) implants were mobile within 8 weeks. Five of the lost implants were single-tooth replacements, of which 2 were immediate placements into extraction sockets.

Chaushu and colleagues²⁷ studied a group of 26 immediately restored cylindrical, press-fit hydroxyapatite-coated implants. Seventeen implants were placed in immediate extraction sockets and 9 were placed in healed alveolar ridges. Occlusal contact in centric occlusion was described as “minimized.” Three of 17 implants placed in extraction sockets failed within the first month, for a survival rate of 82.4%, while all of the implants placed in healed ridges survived. All of the failed implants were placed in the maxilla using a combination of conventional drilling and osteotome bone compression for site preparation. It is important to note that this is the only publication reviewed in this section where press-fit, cylindrical implants were evaluated for immediate restoration.

Following up on their earlier work, Malo and associates²⁸ coordinated a multicenter study with 116 machined Brånemark System implants (Nobel Biocare) with various diameters and configurations placed in 76 patients. These implants were placed in the esthetic zone using surgical techniques of underpreparation of the apical osteotomies to increase initial stability such that insertion torque was greater than 30 Ncm for all implants. Twenty-four patients in this group smoked more than 10 cigarettes per day. The authors reported a 96.5% (112 of 116) success rate for integration and 100% (22 of 22) integration in fresh extraction sockets. None of the smokers lost implants, leading the authors to conclude that initial implant stability was more important than smoking in influencing implant survival and normal healing with this group. A higher failure rate was noted with 3.3-mm-diameter implants, although this was not statistically significant because of the small sample size.

Glauser and coworkers²⁹ reported on a 38-patient series in which 102 Brånemark System Mk IV TiUnite implants were placed (Nobel Biocare); 23 were placed in immediate extraction sites and immediately loaded, 8 were placed in incompletely healed extraction sites, and 71 were placed in healed sites. Twelve smokers were included. Ninety-seven percent (99 of 102) of the implants were clinically successful at 12 months. The authors concluded that neither smoking nor immediate or recent extraction sites had an effect on survival outcome. One patient, who accounted for all of the failed implants, developed an early postoperative infection from a simultaneous guided bone regeneration procedure.

Table 3 Publications on Immediate Restorations in Extraction Sockets

| Authors | Implant system | No. of implants integrated/ placed | Success rate (%) | No. of integrated/placed control implants |
|-------------------------------------|--------------------------|---------------------------------------|------------------|---|
| Wöhrlé 1998 ¹⁵ | Steri-Oss TPS and HA | 14/14 | 100.0 | |
| Malo et al 2000 ²⁵ | Brånemark MKII machined | 23/27 | 85.2 | |
| Hui et al 2001 ²³ | Brånemark machined | 13/13 | 100.0 | 11/11 unloaded in sockets |
| Chausu et al 2001 ²⁷ | Various HA cylindrical | 14/17 | 82.4 | 9/9 in healed ridges |
| Malo et al 2003 ²⁸ | Brånemark machined | 22/22 | 100.0 | |
| Glauser et al 2003 ²⁹ | Brånemark TiUnite | 23/23 | 100.0 | |
| Cannizzaro/Leone 2003 ²¹ | Centerpulse Spline Twist | 46/46 | 100.0 | 45/46 unloaded in sockets |
| Kan et al 2003 ²⁰ | Steri-Oss HA | 35/35 | 100.0 | |
| Overall | | 190/197 | 96.4 | |

Degidi and Piattelli⁵ followed 646 implants under various clinical conditions. While they did not specifically report statistics of extraction sockets versus healed ridges, they indicated that they only had 2 failures with 58 single-tooth implants. Both of these failures occurred in immediate extraction cases where bone condensation was performed for site preparation. The Frialit-2 implants (Friadent, Mannheim, Germany) used in these cases had few macro-geometric features to enhance primary stability. In addition, the authors noted that in both cases the patients exhibited parafunctional habits that applied excessive forces to the implants early in the healing process.

Eight publications that lent themselves to summation and comparison are shown in Table 3. The data pooled from subsets of patients indicate that 197 implants were placed into extraction sockets, resulting in clinical integration of 190, for a clinical success rate of 96.4%. Comparison of the results and conclusions of some articles indicate that a few authors documented poorer integration rates in immediate placement situations. Those authors who achieved high success rates in either condition include Wöhrlé,¹⁵ Hui and coworkers,²³ Glauser and associates,²⁹ and Malo and colleagues,²⁸ who reported common strategies to optimize results. They favored implants with macro-geometric features such as threads to increase immediate bone-to-implant stability and contact. Surgical procedures were modified to increase apical bone density, including underdrilling and self-tapping. Occlusal loads were reduced, with provisional restorations left out of occlusion. Cigarette smoking did not appear to be a factor in achieving integration. Circumstantial reports suggest that implant site preparation through bone condensation may not be optimal for immediate restoration in extraction sockets, in comparison to early loading applications where this type of bone preparation did not appear to affect outcomes.

Single Teeth Versus Multiple Splinted Teeth

Many authors have demonstrated high success rates with immediately restored implants in partial-arch configurations. Case reports and case-control series demonstrating nearly 100% success rates have been reported by Malo and coworkers,^{25,28} Jaffin and associates,³⁰ Chatzistavrou and coworkers,³¹ Degidi and Piattelli,⁵ Calandriello and associates,³² and others. Degidi and Piattelli⁵ reported 100% success of implants (166 of 166) supporting fixed partial dentures in their nonloaded groups. Rocci and colleagues²⁶ reported significantly higher integration rates with multiple-tooth conditions (94%) than with single teeth (81%) using machined-surface implants.

Implant Surface

In a different patient series comparing the influence of implant surface on clinical results, Rocci and coworkers³³ reported a 95.5% success rate with Nobel Biocare TiUnite (roughened-surface) implants in 2- to 4-unit splints but an 85.5% success rate with machined-surface Nobel Biocare implants. This difference in success rate was more pronounced when evaluating implants placed into type 4 bone, where 45% (5 of 11) of machined-surface implants failed and only 8% (1 of 12) of roughened-surface implants failed. These findings are similar to those of Glauser and associates,^{6,29} who demonstrated poor success with machined-surface implants in poor bone quality but good success when roughened surfaces were used in these areas.

Number of Implants, Occlusion, and Placement Technique

One strategy used to enhance success rates has been to increase the number of implants. Calandriello and colleagues³² used 1 implant per tooth and obtained 98% survival. Degidi and Piattelli⁵ recommended a prosthetic unit-to-implant ratio of at least 1.4 in the maxilla and 1.5 in the mandible. They

further recommended that restorations be fabricated out of occlusion, in agreement with Malo and coworkers,²⁸ while Calandriello and colleagues³² recommended light occlusal contact in centric occlusion. In an early loading study, where most restorations were placed within 1.5 weeks of implant surgery, Bogaerde and associates¹² similarly reported high success with light occlusal contact. Most authors recommend altering implant surgical procedures to increase initial stabilization by avoiding tapping the osteotomy sites and by underdrilling the apical width of the osteotomies to increase apical compression. This was specifically mentioned by Malo and coworkers,²⁵ Calandriello and colleagues,³² and Bogaerde and associates.¹²

Bone Density and Quality

Numerous references have been made in the preceding sections regarding the impact of, or association between, bone density or quality and implant success with immediately restored or loaded implants. This implicit relationship between bone density, initial implant stability, and successful osseointegration has been generally accepted by clinicians and confirmed in the literature in relation to conventional loading protocols as described by Jaffin and Berman.³⁴ Mirroring these findings in immediate restoration and loading conditions, Rocci and coworkers²⁶ noted survival of 22 of 27 (81%) machined titanium implants placed in "soft bone" but 66 of 70 (94%) in dense bone, which was statistically significant at $P > .02$. Similarly, Glauser and associates⁶ noted the survival of 66% of implants placed in type 4 bone, but 91% in all other types of bone. In reporting 100% success for integration, Cannizzaro and Leone²¹ noted that 38 of 48 of their immediately loaded implants were placed in bone density type 2⁸ or denser. None of the implants in their study were placed in bone density of less than type 3.

Several authors refer to their stability criteria for immediately loading or restoring dental implants, regardless of bone quality. Wöhrle¹⁵ sought insertion torque of 45 Ncm for single restorations and Hui and associates²³ indicated the need for 40 to 50 Ncm. Horiuchi and coworkers³⁵ observed a mean insertion torque of 42 Ncm for implants used in mandibular full-arch immediate loading cases. Bogaerde and associates¹² recommended a minimum insertion torque of 40 Ncm. Calandriello and coworkers³² indicated that their requirements for immediate loading were a minimum insertion torque of 60 Ncm for single teeth, 45 Ncm for implants supporting partial-arch restorations, and 32 Ncm for implants supporting full-arch restorations. Andersen and coworkers,¹⁶ Malo and associates,^{25,28} Degidi

and Piattelli,⁵ and Lorenzoni and colleagues¹⁹ all indicated that their minimum insertion torque values were 30 to 35 Ncm. Glauser and coworkers²⁹ reported a mean insertion torque of 27 Ncm in their later study. Although Cannizzaro and Leone²¹ did not report insertion torque value, they reported that abutments were torqued to 30 Ncm, indicating that the implants achieved at least this degree of stability.

A few authors have begun to include resonance frequency analysis data^{36,37} in assessing implant stability. Calandriello and coworkers²⁴ reported a mean implant stability quotient (ISQ) of 76 and a minimum of 58 at implant placement for their single molar implant study using 5-mm-diameter implants. Glauser and associates²⁹ reported a mean ISQ at placement of 71 (SD = 8). Of particular interest in this study was the observation of a rapid decrease in mean ISQ value to 63 at 1 week, which gradually increased toward baseline during the 1-year observation period.

Implants with high initial stability appear to survive well under immediate restoration or loading protocols. It would seem that implants placed in softer bone are less stable than those placed in denser bone unless surgical strategies to increase stability are applied. Studies that use insertion torque values are in general agreement that the values should be at least 30 to 35 Ncm. Resonance frequency analysis may prove to be another useful method to aid in selection of the loading protocol.

Implant Surface/Geometry

Most of the data presented on immediately restored and loaded implants have been collected from studies with threaded implants. An exception is the report by Chaushu and associates,²⁷ who used press-fit cylinders and reported a relatively high failure rate (17.6%) in extraction sockets but a 100% success rate in healed ridges.

Implants with a sparse thread pattern have also been evaluated. Degidi and Piattelli⁵ used 82 Frialit-2 implants with immediate loading and reported 6 failures, for a success rate of 93.7%. In immediately restored, unloaded implant sites, they reported 2 failures of 62 Frialit-2 implants (Friadent) in extraction sockets, for a success rate of 96.6%. The 2 failures in this group occurred in extraction sockets that were prepared using bone condensation procedures. Lorenzoni and coworkers¹⁹ reported on immediate restoration of Frialit-2 implants in extraction sockets and had 100% success using occlusal splints for 2 months after placement to eliminate forces on the implants.

Few investigators have directly compared the integration rates of roughened, threaded surfaces with those of machined, threaded surfaces. Rocci

and coworkers³³ noted a significant increase in success rate when comparing Nobel Biocare TiUnite surface threaded implants with machined-surface threaded implants. They found the success rate for the roughened-surface implants to be 95.5% (63 of 66) versus 85.5% (47 of 55) for the machined threaded surfaces. The difference in success rate was particularly striking when evaluating implants placed in poor quality, type 4 bone⁸: 1 of 12 rough-surfaced implants failed, compared with 5 of 11 machined-surface implants. These results are similar to those in case series reported by Glauser and associates.^{6,29}

Alternatively shaped implants have been developed specifically for immediate restoration and loading applications. These include the Altiva NTR System (Altiva, Minneapolis, MN) and the Sargon system (Sargon Enterprises, Beverly Hills, CA). Buchs and coworkers³⁸ reported on the Altiva NTR, which is a 1-piece implant with a dual helical thread pattern designed to increase initial bone stability and eliminate prosthetic abutments and screws. The authors reported on a series of 142 implants that were used in single-tooth (51) and partial-arch (91) clinical situations. Smokers and unhealed extraction sockets were excluded from the study. Nine failures were reported, for a success rate of 93.7%. When the data were further refined, they noted success rates of 83.3% (10 of 12) in type 1 bone, 95.7% (45 of 47) in type 2 bone, 88.9% (24 of 27) in type 3 bone, and 71.4% (5 of 7) in type 4 bone.⁸ One hundred twenty-six of 142 implants were followed at least 1 year, and the implant survival rate did not change after the second month. No data are available to address success criteria of bone level stability, radiographic changes, or gingival indices for this implant system.

Jo and coworkers³⁹ reported on 286 expandable implants manufactured by Sargon and used in 75 patients. Eighty-two of 90 implants placed in extraction sockets were immediately loaded, and 164 of 196 implants placed into healed ridges were immediately loaded. The unloaded implants, including all implants placed into type 4 bone, were deemed not sufficiently stable for immediate loading and were allowed to heal conventionally. The implants were designed to allow for expansion of the apical wings of the implant to re-establish intimate contact with the surrounding bone if implant mobility was noted in the first few weeks of healing. Two hundred eight of 286 implants required apical expansion during early healing. Results indicated that 81 of 82 (98.8%) immediately restored implants placed in extraction sockets and 156 of 164 (95.1%) immediately restored implants placed in

healed ridges survived a minimum of 13 months and up to the maximum observation period of 40 months. Sixty-nine implants had 75 complications. Thirteen of these implants failed. Some of the complications appeared to be unique to the implant design. Implant fracture was not observed. As with the Altiva NTR system, long-term success criteria and data are not available for this implant system.

In summary, the available literature, which is available mostly in a case-series format, is limited and inconclusive with regard to surface and shape characteristics for implants used for immediate restoration and loading. A strong inference can be drawn that implants with increased macroscopic stabilization features such as threads and microscopic enhancements such as surface treatments appear to have improved integration rates compared with smoother designs. TPS and hydroxyapatite coatings, SLA, and increased oxidation appear to improve integration success over machined surfaces, particularly in areas of challenging bone quality. Newer designs like mechanical expansion and large helical threads may offer alternative methods of initial stabilization, but it has not yet been shown that implants with these alternative geometries can achieve long-term success as currently defined.

Stability of Results

It is clear that immediately restored and loaded implants in partial- and single-tooth applications can achieve integration using many implant systems and protocols. Other clinical outcomes that have been evaluated include hard and soft tissue changes. Ericsson and coworkers²² noted that once implants were restored, they lost a mean of 0.1 mm of bone over the 1-year evaluation period, which was similar to data obtained from their control group of delayed loaded implants. Lorenzoni and associates¹⁹ noted that implants placed with an immediate restoration demonstrated 0.45 mm mesial resorption and 0.75 mm distal crestal resorption at 6 and 12 months, which was less than that observed for a standard 2-stage approach. Hui and coworkers²³ reported crestal bone loss of no more than 0.6 mm during their 16-month observation period. Andersen and associates¹⁶ noted a mean gain of radiographic crestal bone level of 0.53 mm during their 5-year observation period. This gain was explained as closure of the vertical defects of the extraction socket walls toward the implant surface. Cannizzaro and Leone²¹ noted radiographic bone loss of 0 to 1 mm on 95.7% (44 of 46) of their immediately loaded implants and 93.3% (43 of 44) of their control group at 24 months, indicating no statistical difference between the immediately loaded and conventional treatment modalities.

Degidi and Piattelli⁵ reported mean bone loss of 1.1 mm after 5 years on 87 immediately restored or loaded implants. Rocci and associates²⁶ noted mean bone loss of 1.0 mm after 1 year, 0.4 mm during the second year, and 0.1 mm in the third year of their study. In their single-tooth, mandibular molar study, Rocci and colleagues³³ similarly measured a mean of 0.9 mm crestal bone loss with TiUnite (ie, rough-surfaced) implants and 1.0 mm with machined-surface implants. Malo and coworkers²⁸ found mean bone loss of 1.1 mm, Calandriello and associates³² measured mean bone loss of 1.2 mm, and Glauser and colleagues²⁹ measured mean bone loss of 1.2 mm in their studies after 1 year. These bone loss measurement data are similar to those reported for conventional loading protocols.^{40,41} Crestal bone resorption data were not reported for the less traditional implant designs of Sargon and Altiva NTR.

Of particular interest to clinicians when placing implants in the esthetic zone are the stability and behavior of soft tissue contours. Achieving stable osseointegration is an important element of predictable implant dentistry, but preserving or creating stable, harmonious soft tissue contours is also of paramount importance. Wöhrle,¹⁵ Hui and coworkers,²³ and Kan and associates²⁰ reported gingival marginal changes of immediately restored implants. Wöhrle¹⁵ noted minimal marginal tissue changes in 12 of 14 patients and recession of 1 to 1.5 mm in the remaining 2 implants. Hui and coworkers²³ did not report data on soft tissue stability, but noted that the esthetic results in their immediately restored sites were superior to those achieved with a staged approach because of gingival architecture preservation.

In a carefully documented study, Kan and associates²⁰ followed 35 maxillary anterior immediately restored implants placed into extraction sockets. After 1 year, they noted radiographic crestal bone loss of 0.26 mm mesially and 0.22 mm distally. Gingival marginal recession was 0.55 mm midfacially, 0.53 mm at the mesial papilla, and 0.39 mm at the distal papilla. These changes are similar to those reported for conventional loading protocols by Bengazi and coworkers⁴² and are slightly less than those reported by Small and Tarnow.⁴³ Additionally, a histomorphometric study in macaque monkeys by Siar and coworkers⁴⁴ supports the observation that no significant differences in crestal bone level or gingival margin location were seen between immediately loaded and conventionally loaded implants.

In general, the case reports and studies indicate that once immediately loaded implants integrate, they appear to have longitudinal bone loss and soft tissue stability comparable to those of conventionally loaded implants. Limited data suggest that immediate

restoration of implants in the esthetic zone might facilitate and stabilize gingival architecture more than a staged approach. No evidence suggests that deleterious gingival complications can be directly attributed to immediate restoration or loading protocols.

REFERENCES

1. U.S. Department of Health and Human Services. Oral Health in America: A Report of the Surgeon General. Rockville, MD: U.S. Department of Health and Human Services, National Institutes of Health, National Institute of Dental and Craniofacial Research, 2000. NIH publication 00-4713.
2. Brunski JB. In vivo bone response to biomechanical loading at the bone/dental-implant interface. *Adv Dent Res* 1999;13:99-119.
3. Cameron HU, Pilliar RM, MacNab I. The effect of movement on the bonding of porous metal to bone. *J Biomed Mater Res* 1973;7:301-311.
4. Szmukler-Moncler S, Piattelli A, Favero GA, Dubruille J-H. Considerations preliminary to the application of early and immediate loading protocols in dental implantology. *Clin Oral Implants Res* 2000;11:12-25.
5. Degidi M, Piattelli A. Immediate functional and non-functional loading of dental implants: A 2- to 60-month follow-up study of 646 titanium implants. *J Periodontol* 2003;74:225-241.
6. Glauser R, Rée A, Lundgren A, Gottlow J, Hämmerle CHF, Schärer P. Immediate occlusal loading of Brånemark implants applied in various jawbone regions: A prospective, 1-year clinical study. *Clin Implant Dent Relat Res* 2001;3:204-213.
7. Cochran DL, Buser D, ten Bruggenkate CM, et al. The use of reduced healing times on ITI implants with a sandblasted and acid-etched (SLA) surface: Early results from clinical trials on ITI SLA implants. *Clin Oral Implants Res* 2002;13:144-153.
8. Lekholm U, Zarb GA. Patient selection and preparation. In: Brånemark P-I, Zarb GA, Albrektsson T (eds). *Tissue-Integrated Prosthesis: Osseointegration in Clinical Dentistry*. Chicago: Quintessence, 1985:199-209.
9. Rocuzzo M, Bunino M, Prioglio F, Bianchi SD. Early loading of sandblasted and acid-etched (SLA) implants: A prospective split-mouth comparative study. *Clin Oral Implants Res* 2001;12:572-578.
10. Rocuzzo M, Wilson T. A prospective study evaluating a protocol for 6 weeks' loading of SLA implants in the posterior maxilla: One-year results. *Clin Oral Implants Res* 2002;13:502-507.
11. Testori T, Del Fabro M, Feldman S, et al. A multicenter prospective evaluation of 2-month loaded Osseotite implants placed in the posterior jaws: 3-year follow-up results. *Clin Oral Implants Res* 2002;13:154-161.
12. Bogaerde LV, Pedretti G, Dellacassa P, Mozzati M, Rangert B. Early function of splinted implants in maxillas and posterior mandibles using Brånemark System machined-surface implants: An 18-month prospective clinical multicenter study. *Clin Implant Dent Relat Res* 2003;5(suppl 1):21-28.
13. Cooper L, Felton DA, Kugelberg CF, et al. A multicenter 12-month evaluation of single-tooth implants restored 3 weeks after 1-stage surgery. *Int J Oral Maxillofac Implants* 2001;16:182-192.

14. Kupeyan HK, May KB. Implant and provisional crown placement: A one-stage protocol. *Implant Dent* 1998;7: 213–217.
15. Wöhrle PS. Single-tooth replacement in the aesthetic zone with immediate provisionalization: Fourteen consecutive case reports. *Pract Periodontics Aesthet Dent* 1998;10: 1107–1114.
16. Andersen E, Haanæs HR, Knutsen BM. Immediate loading of single-tooth ITI implants in the anterior maxilla: A prospective 5-year pilot study. *Clin Oral Implants Res* 2002;13: 281–287.
17. Aires I, Berger J. Immediate placement in extraction sites followed by immediate loading: A pilot study and case presentation. *Implant Dent* 2002;11:87–94.
18. Touati B, Guez G. Immediate implantation with provisionalization: From literature to clinical implications. *Pract Periodontics Aesthet Dent* 2002;14:699–707.
19. Lorenzoni M, Pertl C, Wimmer G, Wegscheider WA. Immediate loading of single-tooth implants in the anterior maxilla. Preliminary results after one year. *Clin Oral Implants Res* 2003;14:180–187.
20. Kan YK, Rungcharassaeng K, Lozada J. Immediate placement and provisionalization of maxillary anterior single implants: 1-year prospective study. *Int J Oral Maxillofac Implants* 2003;18:31–39.
21. Cannizzaro G, Leone M. Restoration of partially edentulous patients using dental implants with a microtextured surface: A prospective comparison of delayed and immediate full occlusal loading. *Int J Oral Maxillofac Implants* 2003;18: 512–522.
22. Ericsson I, Nilson H, Nilner K, Randow K. Immediate functional loading of Brånemark single tooth implants: An 18-month clinical pilot follow-up study. *Clin Oral Implants Res* 2000;11:26–33.
23. Hui E, Chow J, Li D, Liu J, Wat P, Law H. Immediate provisional for single-tooth implant replacement with Brånemark System: Preliminary report. *Clin Implant Dent Relat Res* 2001;3:79–86.
24. Calandriello R, Tomatis M, Vallone R, Rangert B, Gottlow J. Immediate occlusal loading of single lower molars using Brånemark System wide-platform TiUnite implants: An interim report of a prospective open-ended clinical multicenter study. *Clin Implant Dent Relat Res* 2003;5(suppl 1): 74–80.
25. Malo P, Rangert B, Dvarsater L. Immediate function of Brånemark implants in the esthetic zone: A retrospective clinical study with 6 months to 4 years of follow-up. *Clin Implant Dent Relat Res* 2000;2:138–146.
26. Rocci A, Martignoni M, Gottlow J. Immediate loading in the maxilla using flapless surgery, implants placed in predetermined positions, and prefabricated provisional restorations: A retrospective 3-year clinical study. *Clin Implant Dent Relat Res* 2003;5(suppl 1):29–36.
27. Chaushu G, Chaushu S, Tzohar A, Dayan D. Immediate loading of single-tooth implants: Immediate versus non-immediate implantation. A clinical report. *Int J Oral Maxillofac Implants* 2001;16:267–272.
28. Malo P, Friberg B, Polizzi G, Gualini F, Vighagen T, Rangert B. Immediate and early function of Brånemark System implants placed in the esthetic zone: A 1-year prospective clinical multicenter study. *Clin Implant Dent Relat Res* 2003;5(suppl 1):37–46.
29. Glauser R, Lundgren AK, Gottlow J, et al. Immediate occlusal loading of Brånemark TiUnite implants placed predominantly in soft bone: 1-year results of a prospective clinical study. *Clin Implant Dent Relat Res* 2003;5(suppl 1): 47–55.
30. Jaffin RA, Kumar A, Berman CL. Immediate loading of implants in partially and fully edentulous jaws: A series of 27 case reports. *J Periodontol* 2000;71:833–838.
31. Chatzistavrou M, Felton DA, Cooper LF. Immediate loading of dental implants in partially edentulous patients: A clinical report. *J Prosthodont* 2003;12:26–29.
32. Calandriello R, Tomatis M, Rangert B. Immediate functional loading of Brånemark System implants with enhanced initial stability: A prospective 1- to 2- year clinical and radiographic study. *Clin Implant Dent Relat Res* 2003;5(suppl 1): 10–20.
33. Rocci A, Martignoni M, Gottlow J. Immediate loading of Brånemark system TiUnite and machined surface implants in the posterior mandible: A randomized open-ended clinical trial. *Clin Implant Dent Relat Res* 2003;5(suppl 1): 57–63.
34. Jaffin RA, Berman CL. The excessive loss of Brånemark fixtures in type IV bone: A 5-year analysis. *J Periodontol* 1991; 62:2–4.
35. Horiuchi K, Uchida H, Yamamoto K, Sugimura M. Immediate loading of Brånemark System implants following placement in edentulous patients: A clinical report. *Int J Oral Maxillofac Implants* 2000;15:824–830.
36. Meredith N. On the Clinical Measurement of Implant Stability and Osseointegration [thesis]. Göteborg: Göteborg University, 1997.
37. Glauser R, Portmann M, Ruhstaller P, Lundgren AK, Hämerle CHF, Gottlow J. Stability measurements of immediately loaded machined and oxidized implants in the posterior maxilla. A comparative clinical study using resonance frequency analysis. *Appl Osseointegration Res* 2001;2:27–29.
38. Buchs AV, Levine L, Moy P. Preliminary report of immediately loaded Altiva Natural Tooth Replacement dental implants. *Clin Implant Dent Relat Res* 2001;3:97–106.
39. Jo HY, Hobo PK, Hobo S. Freestanding and multiunit immediate loading of the expandable implant: An up-to-40-month prospective survival study. *J Prosthet Dent* 2001;85: 148–155.
40. Jemt T, Patterson P. A 3-year follow-up study on single implant treatment. *J Dent* 1993;21:203–208.
41. Adell R, Lekholm U, Rockler B, et al. Marginal tissue reaction at osseointegrated titanium fixtures. A 3-year longitudinal prospective study. *Int J Oral Maxillofac Surg* 1986;15: 39–52.
42. Bengazi F, Wennstrom JL, Lekholm U. Recession of the soft tissue margin at oral implants. A 2-year longitudinal prospective study. *Clin Oral Implants Res* 1996;7:303–310.
43. Small PN, Tarnow DP. Gingival recession around implants: A 1-year prospective study. *Int J Oral Maxillofac Implants* 2000;15:527–532.
44. Siar CH, Toh CG, Romanos G, et al. Peri-implant soft tissue integration of immediately loaded implants in the posterior macaque mandible: A histomorphometric study. *J Periodontol* 2003;74:571–578.